

Certificate of Calibration

Temp C	Channel Number											
	1	2	3	4	5	6	7	8	9	10	11	12
0.0	-0.3	0.0	-0.2	-0.4	-0.3	-0.5	-0.6	-0.5	-0.3	-0.1	-0.2	-0.3
100.0	99.6	100.0	100.0	99.8	99.7	99.5	99.7	99.8	99.8	99.9	100.1	99.8
200.0	199.9	200.0	200.1	199.9	199.7	199.6	199.7	199.8	199.9	199.9	199.9	199.9
300.0	299.9	300.1	300.1	299.9	299.8	299.8	299.8	299.8	300.0	300.1	300.0	299.9
400.0	399.8	400.2	400.2	399.9	399.8	399.9	399.8	399.7	399.9	400.1	400.0	399.9
500.0	499.9	500.0	500.2	499.9	499.7	499.8	500.0	499.8	500.0	500.0	500.3	500.0
600.0	600.0	600.1	600.2	600.0	599.8	599.8	600.0	600.0	600.0	600.1	600.2	600.0
700.0	700.0	700.1	700.5	700.2	700.1	699.9	700.2	700.2	700.2	700.2	700.3	700.2
800.0	799.8	800.2	800.6	800.2	800.0	800.0	800.3	800.0	800.2	800.2	800.3	800.3
900.0	899.9	900.1	900.3	900.2	899.9	899.8	900.1	899.9	900.1	900.2	900.3	900.2
1000.0	1000.0	1000.2	1000.2	1000.2	1000.0	999.7	1000.1	1000.0	1000.1	1000.0	1000.4	1000.2
1100.0	1099.9	1100.3	1100.2	1100.0	1099.9	1099.9	1100.1	1099.9	1100.1	1100.1	1100.3	1100.2
1200.0	1199.8	1200.3	1200.2	1199.9	1199.8	1199.9	1200.1	1199.8	1200.0	1200.1	1200.2	1200.1
1300.0	1300.0	1300.2	1300.2	1300.0	1300.0	1300.0	1300.1	1300.0	1300.2	1300.1	1300.2	1300.0

Model No:	ChartScan 1400 S/N 155495+155428	McGoff-Bethune, Inc.
Accuracy Required:	+0.6 DegC	5970 Unity Drive
Standard Used:	PA39 S/N 0945-0003	Suite A
Standard Expiration Date:	11-18-99	Norcross, GA 30071
Date of Calibration	16 June 1999	Tel: 770-840-9811
Calibration Expiration:	15 June 2000	Fax: 770-840-7514
Technician:	Lee Bethune	Email: sales@mcgoff-bethune.com
Location	McGoff-Bethune, Inc., Norcross, GA	

Certificate of Calibration

Temp C	13	14	15	16	Average	Max	Min	MaxErr	MinErr
0.0	-0.3	-0.4	-0.5	-0.6	-0.3	0.0	-0.6	0.0	0.6
100.0	99.8	99.8	99.8	99.7	99.8	100.1	99.5	0.1	0.5
200.0	199.9	199.7	199.9	199.7	199.8	200.1	199.6	0.1	0.4
300.0	300.1	300.0	299.8	299.7	299.9	300.1	299.7	0.1	0.3
400.0	400.0	400.1	399.9	399.7	399.9	400.2	399.7	0.2	0.3
500.0	500.0	500.0	500.1	499.7	500.0	500.3	499.7	0.3	0.3
600.0	600.0	600.1	600.0	599.7	600.0	600.2	599.7	0.2	0.3
700.0	700.2	700.1	700.2	700.0	700.2	700.5	699.9	0.5	0.1
800.0	800.1	800.1	800.1	800.0	800.2	800.6	799.8	0.6	0.2
900.0	900.1	900.0	900.1	899.9	900.1	900.3	899.8	0.3	0.2
1000.0	1000.0	1000.0	1000.1	999.8	1000.1	1000.4	999.7	0.4	0.3
1100.0	1100.0	1099.8	1100.1	1099.8	1100.0	1100.3	1099.8	0.3	0.2
1200.0	1200.2	1200.0	1200.2	1199.8	1200.0	1200.3	1199.8	0.3	0.2
1300.0	1300.1	1300.1	1300.0	1299.7	1300.1	1300.2	1299.7	0.2	0.3